

## Passive shielding of the RED-100 detector in an experiment to study the CENNS process

*Thursday, 19 November 2020 14:00 (15 minutes)*

A combined passive shielding of the RED-100 two-phase emission detector has been developed and created, which provides protection from external gamma background and neutron background. The shield is a 5 cm thick copper layer (the inner layer is near the detector) and a water layer with a total thickness of  $\sim 70$  cm. The shielding efficiency has been obtained via the Monte Carlo simulation. Also experimental verification in laboratory conditions using a NaI(Tl) scintillation detector has been done. So the calculated gamma-background attenuation coefficient for the copper shielding, and the complete shielding has been estimated at both approach.

**Primary authors:** Mr LUKYASHIN, Anton (MEPhI / MSU); ON BEHALF OF THE RED COLLABORATION

**Presenter:** Mr LUKYASHIN, Anton (MEPhI / MSU)

**Session Classification:** Приборы и методы экспериментальной ядерной физики

**Track Classification:** Приборы и методы экспериментальной ядерной физики